



AWS, Inc.

910 Bern Ct Suite 100
San Jose, CA 95112
408.453.5070 ph
408.453.5199 fax

Date 13. März 2011

Invoice Number

PO Number

Bin Number

Certification

Ship To:

Company

ATTN:

Address

City

State

Zip code

Phone

Sold To:

Company

ATTN:

Address

City

State

Zip code

Phone

Tool Information

Model: 5132ST

Make: Hazet

Cal Date:

Range: 300

Units: Nm

CW Tolerance %: 4

Due Date:

Serial Number: 2021030

CCW Tolerance %: 6

Calibration Test Report

Wrench Setting	Allowable Deviation	As Found Readings						Largest Deviation	In Spec?
+60	2.4	60.2	60.3	60.4	60.1	60.5	60.3	0.5	YES
+120	4.8								
+180	7.2								
+240	9.6								
+300	12								
-60	3.6								
-120	7.2								
-180	10.8								
-240	14.4								
-300	18								

Traceability Information

Torque Display: _____ Serial Number: _____ Due: _____
Torque Transducer: _____ Serial Number: _____ Due: _____
NIST Numbers: _____
Procedure: _____ Temperature: 22 Humidity: _____
Comments: _____
Technician: Andreas Lang
Signature: _____ Date: 13. März 2011